REVEAL
IMAGING TECHNOLOGIES, INC.

THIS DOCUMENT CONTAINS CONFIDENTIAL INFORMATION PROPRIETARY TO REVEAL IMAGING TECHNOLOGIES INC. IT MUST NOT BE REPRODUCED OR DISCLOSED TO OTHERS OR USED IN ANY OTHER WAY, IN WHOLE OR IN PART, EXCEPT AS AUTHORIZED IN WRITING BY REVEAL IMAGING TECHNOLOGIES, INC.

DOC NO: 40009 Rev: B

TITLE: CT-80 RADIATION SURVEY MAP
RADIATION LIMIT: < 500uR/HR
SYSTEM NO: 040049 (A-5)
DATE: 11/6/09





TEST RECORD

ZONE	RADIATION LEVEL VALUE (< 500 µR/h)	ZONE	RADIATION LEVEL VALUE (< 500 µR/h)
1) ENTRANCE END	108 uR/h	5) EXIT END	88 uR/h
2) RIGHT SIDE ENTRANCE TUNNEL	17 uR/h	6) LEFT SIDE EXIT TUNNEL	22 uR/h
3) RIGHT SIDE GANTRY	24 uR/h	7) LEFT SIDE GANTRY	24 uR/h
4) RIGHT SIDE EXIT TUNNEL	16 uR/h	8) LEFT SIDE ENTRANCE TUNNEL	19 uR/h

Technician Name: [redacted]

Notes: none

w/0-2691613

Instrument: INDUSTRON 451

Rit #:

Serial #: 0606

Cal Due Date: 06/17/10

159.2 KV
2.0 ma