REVEAL IMAGING TECHNOLOGIES, INC.

THIS DOCUMENT CONTAINS CONFIDENTIAL INFORMATION PROPRIETARY TO REVEAL IMAGING TECHNOLOGIES INC. IT MUST NOT BE REPRODUCED OR DISCLOSED TO OTHERS OR USED IN ANY OTHER WAY, IN WHOLE OR IN PART, EXCEPT AS AUTHORIZED IN WRITING BY REVEAL IMAGING TECHNOLOGIES, INC

DOC NO: 40009 Rev: B

TITLE: CT-80 RADIATION SURVEY MAP
RADIATION LIMIT: < 500uR/HR

SYSTEM NO: 040577

DATE: 12/30/09





TEST RECORD

ZONE	RADIATION LEVEL VALUE (< 500 µR/h)	ZONE	RADIATION LEVEL VALUE (< 500 µR/h)
1) ENTRANCE END	365	5) EXIT END	327
2) RIGHT SIDE ENTRANCE TUNNEL	30	6) LEFT SIDE EXIT TUNNEL	69
3) RIGHT SIDE GANTRY	89	7) LEFT SIDE GANTRY	74
4) RIGHT SIDE EXIT TUNNEL	71	8) LEFT SIDE ENTRANCE TUNNEL	43

Technician Name: [redacted]

Notes:

Instrument: Thermo Scientific Radeye PRD

Rit #: 0120

Serial #: 013756

Cal Due Date: 10/26/10