

# REVEAL


DO NOT USE IN SENSITIVE AREA. THIS DOCUMENT CONTAINS INFORMATION WHICH IS THE PROPERTY OF THE GOVERNMENT OF THE UNITED STATES OF AMERICA AND IS SUBJECT TO RESTRICTIONS ON DISCLOSURE AND USE. IT IS TO BE KEPT SECURE AND IS NOT TO BE SHOWN TO UNAUTHORIZED PERSONS.

DOC NO: 40009 Rev: C


TITLE: RADIATION SURVEY MAP FOR  
CT-80, CT-80XL, CT-80XR, B, CT-80DXL  
RADIAITON LIMIT: < 5000 uR/H

DO NOT USE IN SENSITIVE AREA. THIS DOCUMENT CONTAINS INFORMATION WHICH IS THE PROPERTY OF THE GOVERNMENT OF THE UNITED STATES OF AMERICA AND IS SUBJECT TO RESTRICTIONS ON DISCLOSURE AND USE. IT IS TO BE KEPT SECURE AND IS NOT TO BE SHOWN TO UNAUTHORIZED PERSONS.

CT-80 / CT-80XR




CT-80XL / CT-80DXL



SYSTEM NO: 40-576

DATE: 12-07-2010



## TEST RECORD

ZONE	RADIATION LEVEL VALUE (< 500 $\mu$ R/H)	ZONE	RADIATION LEVEL VALUE (< 500 $\mu$ R/H)
1) ENTRANCE END	320	5) EXIT END	122
2) RIGHT SIDE ENTRANCE TUNNEL	155	6) LEFT SIDE EXIT TUNNEL	55
3) RIGHT SIDE GANTRY	N	7) LEFT SIDE GANTRY	54
4) RIGHT SIDE EXIT TUNNEL	88	8) LEFT SIDE ENTRANCE TUNNEL	38
9) RIGHT SIDE XL ENTRANCE		11) LEFT SIDE XL EXIT	
10) RIGHT SIDE XL EXIT		12) LEFT SIDE XL ENTRANCE	

Technician Name:

Name: \_\_\_\_\_

Instrument: 42506/31

Ref #: 205

Serial #: 14320

Cal Due Date: